

بۆ گشت كۆمپانیاكانى ئامپىرى پزىشكى

الى كافة الشركات الاجهزة الطبية

م/پیشکەشکردنى ئۆفەر

عروض للاجهزة الطبية

تکایه ئۆفهرى خۇتان پیشکەش بکەن بۆ ئەم ئامپىره پزىشکیانهى خوارده ، و ماوهى و مرهنتى له سى
سال بهرمو سهروهو بىت .

يرجى تقديم عروضكم للاجهزة الطبية الواردة في الجدول ادناه مع كافة المواصفات المتعلقة بالجهاز مع
تقديم ضمان لمدة ثلاث سنوات فما فوق.

No.	Name of equipment	Qua.
1	YAG SLT laser (2 lasers on 1 platform) 1-Energy: from 0.5mj to at least 10mj (for the Yag laser) 2-Slit lamp with electric table: a- One platform combining both ND: YAG laser &SLT laser on the same table and slit lamp. The slit lamp should have at least 3 magnifications. b-the table: easy access electric table design for wheel chaired patients and the table should be mobile 3-Protective goggles for 1064nm wavelength	1
2	OCT 1-should have both anterior (corneal) and posterior (retinal) modules with accessories need for this purpose 2-should have protocols for analysis of both macula and the optic nerve head 3-should be using spectral domain technology or newer technologies 4-easy access design of its electric table for wheel chaired patients and the table should be mobile type 5-DICOM compatible	1
3	Autorefracto- kerato- tonometer 1- three in one device 2- thermal printer integrated with auto cut 3- DICOM compatible 4- included accessories: electric table	5
4	Trial box and frame 1- Complete set up to 20D sphere and at least 6D astigmatism both concave and convex lenses 2- Prism: commonest range of powers 3- All accessories of refraction: pinhole, zero power lenses, red green trials, Jackson cross cylinders, etc. 4- At least one adult (universal) frame and one pediatric (fixed or variable PD) frame.	6
5	AB scan It should include: 1-A scan(biometry) and IOL calculation: all the formulae of IOL calculation should be available 2-B scan with ability to do biometry in this mode 3-pachymetry 4-A4 printer (black & white) 5- DICOM and EMR compatible	1
6	Ophthalmic operative microscope 1-motorized zoom ,motorized focusing with apochromatic lens 2-assistant microscope(integrated or not) controllably independently from main piece	1

	3-XY coupling and motorized :big travel ranges and preferred. 4-slit lamp as accessory. 5-light source :LED or xenon preferred ,back up light source. 6-magnetic breaks for positioning. 7-foot control: multifunction wireless option is preferred. 8-coaxial illumination type. 9-Retro illumination booster systems preferred. 10-type of microscope :professional ophthalmic surgery microscope of floor type designed for both anterior segment and posterior segment surgeries. 11-accessories required: a-full range of accessories required for retinal surgery whether standard or optional ones. b-keratoscope whether integrated or added as optional system. c-full system of digital HD video and image capturing unit and monitor. d-slit lamp	
7	Vision charts screen 1-Flexible working distance: at least from 2.5 m up to 6m 2-Screen should be at least 17 inch large. 3-Include charts: at least one illiterate, one letter, one numeric ,one picture test, as much as orthopedic tests will be given priority ,patient education materials will be given priority. 4-DICOM compatible.	6
8	Auto lens meter: 1-sensor with multiple point measurement technology. 2-automatic lens type detection. 3-prism measurement capability. 4-UV filters measurement capability. 5-PD measurement capability. 6-thermal printer 7-LED and touch screen 8-DICOM compatible	3

تکایه ریزبهندي ئاميركه به گویرهی ئەم راگهیانندنه بێت لهسەر ئۆفەر . وه هەر برگهیهك به تهنیا پیشكهش بكریت به دیناری عیراقی ، وه ئۆفهرهكان پیشكهشی بهشی ئامیری پزشکی دهکریت لهگهڵ نوسینی ناوی ئامیر و ژمارهکهی .

دووا مؤلّهت بۆ پیشكهش کردن ۲۰۱۵/۲/۲۲

یرجی تدوین تسلسل الجهاز بالاعلان على العرض، وتقدیم عرض كل فقرة على حدی، تقدیم الاسعار بالدينار العراقي.

یتم تقدیم العروض مباشرة الى قسم الاجهزة الطبية مع كتابة اسم ورقم الشركة

آخر موعد للتقدیم ۲۰۱۵/۲/۲۲

